


MASS SPECTROMETRY REPORT

28-Dec-2017
12:22:41

Probe: ESI
Cone: 50v
Desovation Temp :350
Capillary:3.00KV
Extractor: 5v
Gas Flow : 350
Scan ES+
1.13e6

171228-28 51 (0.942) Cm (49:52)



Sample Description

Analyzed date: 2016-12-28
Analyst: YU
Sample: A-16
M.W.: 1749.03
Lot. No.: P161225-WD627154

Instrument

Waters ZQ2000
Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

Waters ZQ2000

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H2O/50%ACN