


HPLC REPORT

Sample: M-16 Analyzed date: 2016-12-23
Analyst: FS
Lot. No.: P161219-SW627443
Column: 4.6*250mm, Welch CX-C18 5um
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
Gradient:

	A	B
0.0min	17%	83%
25.0min	42%	58%
25.1min	100%	0%
30.0min		Stop

Volume: 20µl
Wavelength: 220nm
Flow rate: 1.0ml/min



Rank	Time	Conc.	Area
1	8.741	4.317	117437
2	9.158	86.35	2349010
3	9.363	4.912	133604
4	9.795	4.419	120198
Total		100	2720249