


HPLC REPORT

Sample: R-16 Analyzed date: 2016-12-28
 Analyst: LY
 Lot. No.: P161219-YS627081
 Column: 250*4.6mm,Kromasil-C18-5um
 Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
 Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
 Gradient:

	A	B
0.0min	19%	81%
25.0min	44%	56%
25.1min	100%	0%
30.0min	Stop	

Volume: 10µl
 Wavelength: 220nm
 Flow rate: 1.0ml/min



Peak No	Ret Time	Height	Area	Conc.
1	11.222	1113.364	9326.202	0.2219
2	12.607	24605.871	394955.313	9.3956
3	12.903	308387.406	3734377.250	88.8373
4	12.903	8939.684	31887.770	0.7586
5	13.705	618.733	4794.199	0.1140
6	17.650	3492.313	28273.104	0.6726
Total				100.00