


# MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2016-12-25  
 Analyst: Xu  
 Sample: M-16  
 M.W.: 1880.11  
 Lot. No.: P161215-LL627112

Instrument

SHIMADZU LCMS-2010EV  
 Probe: ESI  
 Nebulizer Gas Flow: 1.5L/min  
 CDL: -20.0v  
 CDL Temp.: 250 °C  
 Block Temp.: 200 °C

Probe Bias:

+ 4.5kv  
 Detector: 1.5kv  
 T. Flow: 0.2ml/min  
 B. Conc.: 50% H2O/50% ACN