


MASS SPECTROMETRY REPORT

27-Sep-2017
10:36:00

Probe: ESI
Cone: 50v
Desolvation Temp :350
Capillary:3.00KV
Extractor: 5v
Gas Flow : 350
Scan ES+
2.65e6

170927-18 1186 (21.921)



Sample Description

Analyzed date: 2014-09-27
Analyst: YU
Sample: C-14
M.W.: 1418.55
Lot. No.: P140921-CJ609526

Instrument

Waters ZQ2000
Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

Waters ZQ2000

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H2O/50%ACN