


# MASS SPECTROMETRY REPORT

12-Oct-2017  
08:16:54

Probe: ESI  
Cone: 50v  
Desovation Temp :350  
Capillary:3.00KV  
Extractor: 5v  
Gas Flow : 350  
Scan ES+  
1.44e5

171012-1 527 (9.741) Cm (522:531)



Sample Description

Analyzed date: 2014-10-12  
Analyst: YU  
Sample: C-14  
M.W.: 1603.83  
Lot. No.: P140929-SY608897

Instrument

Waters ZQ2000  
Probe: ESI  
Nebulizer Gas Flow: 1.5L/min  
CDL: -20.0v  
CDL Temp.: 250 °C  
Block Temp.: 200 °C

Probe Bias: + 4.5kv  
Detector: 1.5kv  
T. Flow: 0.2ml/min  
B. Conc.: 50% H<sub>2</sub>O/50% ACN