


MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2014-9-28
Analyst: Xu
Sample: C-14
M.W.: 1767.96
Lot. No.: P140922-YJ608772

Instrument

Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

SHIMADZU LCMS-2010EV

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H₂O/50%ACN