


HPLC REPORT

Sample: C-13 Analyzed date: 2014-09-29
Analyst: FS
Lot. No.: P140922-YJ608782
Column: 4.6*250mm, Welch CX-C18 5um
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
Gradient:

	A	B
0.0min	25%	75%
25.0min	40%	60%
25.1min	100%	0%
30.0min		Stop

Volume: 20µl
Wavelength: 220nm
Flow rate: 1.0ml/min
mV



Rank	Time	Conc.	Area
1	9.133	2.544	726025
2	9.413	94.07	26847613
3	9.733	2.772	791171
4	11.673	0.6093	173864
Total		100	28538673