


# HPLC REPORT

Sample: S-11P  
Analyst: WJ  
Lot. No.: P140705-CJ594033  
Column: Gemini-NX 5 $\mu$  C18 110A, 4.6\*250mm  
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile  
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water  
Gradient:

Analyzed date: 2014-7-13

	A	B
0.0min	20%	80%
25.0min	45%	55%
25.1min	100%	0%
30.0min	Stop	

Volume: 20 $\mu$ l  
Wavelength: 220nm  
Flow rate: 1.0ml/min



Rank	Time	Conc.	Area	Height
1	7.019	0.8455	22053	2721
2	7.246	0.5539	14447	1640
3	9.650	1.0184	26563	7260
4	9.794	95.6824	2495606	281261
5	10.050	1.2992	33885	3944
6	10.335	0.6006	15665	1768
Total		100	2608219	298594