


MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2014-7-13
 Analyst: Xu
 Sample: D-12
 M.W.: 1407.39
 Lot. No.: P140706-YS593871

Instrument

Probe: ESI
 Nebulizer Gas Flow: 1.5L/min
 CDL: -20.0v
 CDL Temp.: 250 °C
 Block Temp.: 200 °C

SHIMADZU LCMS-2010EV

Probe Bias: + 4.5kv
 Detector: 1.5kv
 T. Flow: 0.2ml/min
 B. Conc.: 50% H2O/50% ACN