


HPLC REPORT

Sample: I-16 Analyzed date: 2016-06-14
Analyst: RZG
Lot. No.: P160605-GY659370
Column: SinoChrom ODS-BP, 4.6*250mm, 5µm
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water

Gradient:

	A	B
0.0min	18%	82%
25.0min	43%	57%
25.1min	100%	0%
30.0min	Stop	

Volume: 20µl
Wavelength: 220nm
Flow rate: 1.0ml/min



Rank	Time	Conc	Area
1	12.390	4.947	726975
2	12.711	88.46	12998912
3	13.380	6.591	968527
Total		100	14694414