


MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2014-9-27
 Analyst: Xu
 Sample: F-15
 M.W.: 1709.73
 Lot. No.: P140921-HX608723

Instrument

Probe: ESI
 Nebulizer Gas Flow: 1.5L/min
 CDL: -20.0v
 CDL Temp.: 250 °C
 Block Temp.: 200 °C

SHIMADZU LCMS-2010EV

Probe Bias: + 4.5kv
 Detector: 1.5kv
 T. Flow: 0.2ml/min
 B. Conc.: 50%H₂O/50%ACN