


HPLC REPORT

Sample: G-15
Analyst: HXH
Lot. No.: P140719-SY595198
Column: Gemini-NX 5 μ C18 110A, 4.6*250mm
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
Gradient:

Analyzed date: 2014-07-21

	A	B
0.0min	10%	90%
25.0min	35%	65%
25.1min	100%	0%
30.0min	Stop	

Volume: 20 μ l
Wavelength: 220nm
Flow rate: 1.0ml/min



Rank	Time	Conc.	Area	Height
1	9.619	3.4330	320065	28936
2	10.305	6.0675	565681	42114
3	10.447	88.3087	8233108	842541
4	10.725	2.1908	204251	46487
Total	100		9323105	960078