


MASS SPECTROMETRY REPORT

21-Jul-2017
11:59:20

Probe: ESI
Cone: 50v
Desovation Temp :350
Capillary:3.00KV
Extractor: 5v
Gas Flow : 350
Scan ES+
2.77e6

170721-11 146 (2.698) Cm (145:148)



Sample Description

Analyzed date: 2014-07-21
Analyst: YU
Sample: G-15
M.W.: 1638.80
Lot. No.: P140719-SY595198

Instrument

Waters ZQ2000
Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

Waters ZQ2000

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50% H₂O/50% ACN