




Sample Description

Analyzed date: 2014-7-13
Analyst: Xu
Sample: C-16
M.W.: 1683.83
Lot. No.: P140705-YJ593517

Instrument

Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

SHIMADZU LCMS-2010EV

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H₂O/50%ACN