


HPLC REPORT

Sample: E-14
Analyst: WJ
Lot. No.: P140601-HS546398
Column: Gemini-NX 5 μ C18 110A, 4.6*250mm
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
Gradient:

Analyzed date: 2014-6-9

	A	B
0.0min	15%	85%
25.0min	40%	60%
25.1min	100%	0%
30.0min	Stop	

Volume: 20 μ l
Wavelength: 220nm
Flow rate: 1.0ml/min
mV



Rank	Time	Conc.	Area	Height
1	10.158	1.9341	20711	2292
2	10.387	95.9493	1027430	133421
3	10.679	1.4263	15273	1021
4	11.101	0.6903	7391	1007
Total		100	1070805	137741