


HPLC REPORT

Sample: N-14
Analyst: WJ
Lot. No.: P140601-HS224870
Column: Gemini-NX 5 μ C18 110A, 4.6*250mm
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
Gradient:

Analyzed date: 2014-6-9

	A	B
0.0min	20%	80%
25.0min	45%	55%
25.1min	100%	0%
30.0min	Stop	

Volume: 20 μ l
Wavelength: 220nm
Flow rate: 1.0ml/min



Rank	Time	Conc.	Area	Height
1	9.147	3.2926	238347	15277
2	9.513	95.1759	6889662	664056
3	9.750	0.4993	36146	11165
4	13.778	0.3930	28447	3653
5	14.392	0.6392	46268	4703
Total		100	7238870	698854