


MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2014-6-12
Analyst: SHAO
Sample: N-16
M.W.: 1877.21
Lot. No.: P140607-SY587885

Instrument

Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

SHIMADZU LCMS-2020

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H2O/50%ACN