


HPLC REPORT

Sample: P-16 Analyzed date: 2016-05-18
 Analyst: LY
 Lot. No.: P160509-LL654371
 Column: Kromasil-C18, 4.6*250mm, 5µm
 Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
 Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
 Gradient:

	A	B
0.0min	10%	90%
25.0min	35%	65%
25.1min	100%	0%
30.0min	Stop	

 Volume: 10µl
 Wavelength: 220nm
 Flow rate: 1.0ml/min



Peak No	Ret Time	Height	Area	Conc.
1	9.005	146.915	1215.484	0.0515
2	9.923	519.728	5279.783	0.2236
3	10.383	421.800	8266.809	0.3500
4	10.767	2149.167	22405.840	0.9487
5	10.928	16004.781	2206039.750	93.4104
6	10.928	3615.165	48210.773	2.0414
7	11.985	5256.888	70245.859	2.9744
Total				100.00