


HPLC REPORT

Sample: R-15 Analyzed date: 2016-05-19
 Analyst: FS
 Lot. No.: P160509-LL150628
 Column: 4.6*250mm, Welch CX-C18 5um
 Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
 Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
 Gradient:

	A	B
0.0min	10%	90%
25.0min	35%	65%
25.1min	100%	0%
30.0min	Stop	

Volume: 20µl
 Wavelength: 220nm
 Flow rate: 1.0ml/min



Rank	Time	Conc.	Area
1	17.137	5.119	94971
2	17.655	85.44	1584979
3	17.885	9.027	167473
4	18.473	0.4178	7751
Total		100	1855174