

# HPLC REPORT

Sample: E-16 Analyzed date: 2015-05-18

Analyst: YW

Lot. No.: P150514-SY654930

Column: 4.6×250mm, Kromasil 100-5-C18

Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile


Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water

Gradient:	A	B
0.0min	28%	72%
25.0min	53%	47%
25.1min	100%	0%
30.0min	Stop	

Volume: 5µl

Wavelength: 220nm

Flow rate: 1.0ml/min



Peak	Time	Conc.	Area	Height
1	5.543	0.2189	9782	1483
2	6.649	0.721	32224	2794
3	7.381	1.765	78898	9310
4	7.772	1.157	51709	5463
5	8.381	3.547	158548	13619
6	8.915	91.56	4092140	460197
7	9.165	1.03	46056	5447

Total		100	4469357	498313
-------	--	-----	---------	--------