


MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2015-08-24
 Analyst: YU
 Sample: G-15
 M.W.: 1600.83
 Lot. No.: P150822-SY673214

Instrument

Agilent-6125B
 Probe: ESI
 Nebulizer Gas Flow: 1.5L/min
 CDL: -20.0v
 CDL Temp.: 250 °C
 Block Temp.: 200 °C

Probe Bias: + 4.5kv
 Detector: 1.5kv
 T. Flow: 0.2ml/min
 B. Conc.: 50% H₂O/50% ACN