


# MASS SPECTROMETRY REPORT



## Sample Description

Analyzed date: 2016-05-21  
 Analyst: YU  
 Sample: C-11  
 M.W.: 1138.25  
 Lot. No.: P160514-SY655831

## Instrument

Agilent-6125B  
 Probe: ESI  
 Nebulizer Gas Flow: 1.5L/min  
 CDL: -20.0v  
 CDL Temp.: 250 °C  
 Block Temp.: 200 °C

Probe Bias: + 4.5kv  
 Detector: 1.5kv  
 T. Flow: 0.2ml/min  
 B. Conc.: 50% H<sub>2</sub>O/50% ACN