


MASS SPECTROMETRY REPORT

18-Jan-2018
12:01:37

Probe: ESI
Cone: 50v
Desovation Temp :350
Capillary:3.00KV
Extractor: 5v
Gas Flow : 350
Scan ES+
1.11e7

180118-22 692 (12.790)



Sample Description

Analyzed date: 2016-01-18
Analyst: YU
Sample: L-15
M.W.: 1727.04
Lot. No.: P160116-SY632392

Instrument

Waters ZQ2000
Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

Waters ZQ2000

Probe Bias: + 4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H2O/50% ACN