



Technical Data

Hugh Leifson Medium

M826S

Intended use

Hugh Leifson Medium is used to distinguish between anaerobic and aerobic breakdown of carbohydrate (glucose).

Composition**

Ingredients	Gms / Litre
Peptone	2.000
Sodium chloride	5.000
Dipotassium phosphate	0.300
Glucose (Dextrose)	10.000
Bromothymol blue	0.030
Agar	3.000
Final pH (at 25°C)	7.1±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 20.33 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Dispense into test tubes in duplicate for aerobic and anaerobic fermentation. Sterilize by autoclaving at 10 lbs pressure (115°C) for 20 minutes. Cool the tubed medium in an upright position

Principle And Interpretation

Hugh Leifson Medium was formulated by Hugh and Leifson (1). They described the taxonomic significance of fermentative and oxidative metabolism of carbohydrates in gram-negative intestinal bacteria.

It is recommended by BIS (2) for the isolation and cultivation of *Vibrio cholerae* and other *Vibrio* species which cause food poisoning.

The medium contains a high concentration of carbohydrate and low concentration of peptone to avoid the possibility of an aerobic organism utilizing peptone and producing an alkaline condition which would neutralize slight acidity produced by an oxidative organism (3). Dipotassium phosphate promotes fermentation and acts as pH controlling buffer. Agar concentration enables the determination of motility and aids in distribution of acid throughout the tube produced at the surface of medium. Oxidative organisms produce acid in unsealed tube with little or no growth and no acid formation in sealed tube while fermentative organisms produce acid in both sealed and unsealed tubes.

Type of specimen

Clinical samples - Blood ; Food and dairy samples ; Water samples

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (7,8).

For food and dairy samples, follow appropriate techniques for sample collection and processing as per guidelines (4,5,9).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(6)

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

In Vitro diagnostic Use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets

Limitations :

This medium is general purpose medium and may not support the growth of fastidious organisms.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to bluish green homogeneous free flowing powder

Gelling

Semisolid, comparable with 0.2% Agar gel.

Colour and Clarity of prepared medium

Greenish blue coloured, clear to slightly opalescent gel forms in tubes as butts

Reaction

Reaction of 2.03% w/v aqueous solution at 25°C. pH : 7.1±0.2

pH

6.90-7.30

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-48 hours.

Cultural Response

Organism	Inoculum (CFU)	Motility	Aerobic fermentation	Anaerobic fermentatoin
Cultural Response				
<i>Enterobacter aerogenes</i> ATCC 13048 (00175*)	50-100	positive, growth away from stabline causing turbidity	acid (yellow) and gas production, positive reaction	acid (yellow) and gas production, positive reaction
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	positive, growth away from stabline causing turbidity	acid (yellow) and gas production, positive reaction	acid (yellow) and gas production, positive reaction
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025*)	50-100	positive, growth away from stabline causing turbidity	acid (yellow) production, positive reaction	unchanged (green) or alkaline (blue), negative reaction
<i>Salmonella Typhi</i> ATCC 6539	50-100	positive, growth away from stabline causing turbidity	acid (yellow) and gas production, positive reaction	acid (yellow) and gas production, positive reaction
<i>Shigella sonnei</i> ATCC 25931	50-100	negative, growth along the stabline, surrounding medium	acid (yellow) production, positive reaction	acid (yellow) and gas production, positive reaction

Key- (*) corresponding WDCM folders

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (7,8).

Reference

1. Hugh and Leifson, 1953, J. Bacteriol., 66:24.
2. Bureau of Indian Standards, IS:5887 (Part V) 1976, reaffirmed 1986.
3. MacFaddin J.F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. I, Williams and Co., St. Louis. Wilkins, Baltimore.
4. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
5. Downes F. P. and Ito K., (Ed.), 2001, Compendium of Methods for the Microbiological Examination of Foods, 4th Ed., American Public Health Association, Washington, D.C.
6. Greenberg A. E., Clesceri L. S. and Eaton A. D., (Eds.), 2005, Standard Methods for the Examination of Water and Wastewater, 21st ed., APHA, Washington, D.C.
7. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
8. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S. and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
9. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.


Revision : 02 / 2017



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
B /4-6 , MIDC, Dindori, Nashik MH

www.himedialabs.com



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.